


User's Guide

EXTECH[®]
INSTRUMENTS
A FLIR COMPANY

Humidity & Temperature Datalogger

Model RHT20




Introduction



Congratulations on your purchase of this Temperature and Humidity datalogger. With this meter, you can monitor and log data over long periods of time and then easily transfer the data to a pc for viewing and evaluation. The LCD display provides current or Max/Min temperature, humidity and time information. This meter is shipped fully tested and calibrated and, with proper use, will provide years of reliable service.

Description

- 1 Record LED
- 2 Alarm LED
- 3 LCD Display
- 4 Temperature – Humidity display selection
- 5 MAX/MIN display selection
- 6 Battery Compartment (rear)
- 7 Temperature and Humidity sensors
- 8 USB PC port (bottom)



Display Description

-  Full battery symbol.
-  Weak battery symbol. Replace battery when this appears. Battery life is more than 3 months.

DATE: Current date is displayed

TIME: Current time is displayed
Time and Date automatically swap every 10 sec.

MAX: Maximum value during adatalogging session is displayed

MIN: Minimum value during a datalogging session is displayed


REC: Recording in progress indicator

FULL: Memory is full indicator

%RH: Humidity value is displayed

°C: Celsius temperature units

°F: Fahrenheit temperature units



Operation:


The display and front panel buttons provide a means to view the status of the datalogger, observe the current Temperature or Relative Humidity and to view the MAX and MIN values recorded during a recording session. The datalogger setup conditions such as sample rate, temperature units and alarm values are programmed via the provided software. Refer to the software Help File on the disk for those procedures.

1. Press the MAX/MIN button to view the recorded Max and Min values with the time they were recorded. If, while in this mode, no button is pressed for 40 seconds the display will automatically return to the real time display.
2. The Time and Date display will automatically toggle every 10 seconds
3. When downloading the data to the PC, "-PC-" will be displayed on the LCD. After the download is complete, the datalogger will display the last value recorded, but it will stop recording. The logger needs to be re-set from the software to begin a new recording session.
4. The "REC" LED will flash at the programmed rate when the unit is datalogging.
5. Alarm function: When the measured value exceeds the programmed upper or lower limit and the LED function is selected in the software, the ALM LED will flash one time per minute.
6. If "-LO-" is displayed, the temperature and humidity sensors need attention.
7. The datalogger should be placed upright when in use.
8. If used in low temperature, high humidity environment, the datalogger should be placed upright in a dry area to remove any condensation before downloading the data.
9. Battery life can be extended by leaving the datalogger connected to a pc.
10. Disabling the REC and ALM LED's will extend battery life.

Combination Lock

The logger is supplied with a wall mount case and security combination lock. The lock is shipped with a -0-0-0- code as viewed from the side with the alignment bar. To change the code:

- 1 Press the release to open the lock (1)
- 2 Using a pointed object, press IN and HOLD the locking pin on the bottom of the lock (2).
- 3 Set the new code (3) and release the locking pin



Warranty

EXTECH INSTRUMENTS CORPORATION (A FLIR COMPANY) warrants this instrument to be free of defects in parts and workmanship for **one year** from date of shipment (a six month limited warranty applies to sensors and cables). If it should become necessary to return the instrument for service during or beyond the warranty period, contact the Customer Service Department at (781) 890-7440 ext. 210 for authorization or visit our website www.extech.com for contact information. A Return Authorization (RA) number must be issued before any product is returned to Extech. The sender is responsible for shipping charges, freight, insurance and proper packaging to prevent damage in transit. This warranty does not apply to defects resulting from action of the user such as misuse, improper wiring, operation outside of specification, improper maintenance or repair, or unauthorized modification. Extech specifically disclaims any implied warranties or merchantability or fitness for a specific purpose and will not be liable for any direct, indirect, incidental or consequential damages. Extech's total liability is limited to repair or replacement of the product. The warranty set forth above is inclusive and no other warranty, whether written or oral, is expressed or implied.

Software Installation

1. Place the CD in the drive and then follow the on-screen installation instruction.
2. After the software is installed, keep the CD in the drive and connect the datalogger to the PC by the USB cable.
3. A USB Driver installation window will pop-up. Follow the directions to install the driver.

Software application

The Datalogger USB software is a program for collecting data from the DATA LOGGER when it is connected to a PC or notebook computer. The data may be displayed graphically, as Excel or similar programs. The major functions are all listed in the main window.

System Requirements

Windows 2000 or Windows XP or Vista

Minimum Hardware Requirements:

PC or NoteBook with Pentium 90MHz or higher, 32 MB RAM ;
At least 7 MB byte hard disk space available to install HT Datalogger USB software.
Recommended display resolution 1024X768 with High Color(16 bit).

Software Operation

The Software operation is described in the HELP file .

Calibration and Repair Services

Extech offers repair and calibration services for the products we sell. Extech also provides NIST certification for most products. Call the Customer Care Department for information on calibration services available for this product. Extech recommends that annual calibrations be performed to verify meter performance and accuracy.



Support line (781) 890-7440

Technical Support: Extension 200; E-mail: support@extech.com

Repair & Returns: Extension 210; E-mail: repair@extech.com

Product specifications subject to change without notice

For the latest version of this User Guide, Software updates, and other up-to-the-minute product information, visit our website: www.extech.com
Extech Instruments Corporation, 285 Bear Hill Road, Waltham, MA 02451

Specifications

| | |
|---------------------------|---|
| Display | Multi-function LCD |
| Measurement ranges | 0 to 100% RH -40 to 158°F, -40 to 70°C |
| Resolution | 0.1°, 0.1RH |
| Maximum data points | 16,350 Temperature and 16,350 Humidity values |
| Sample rate | 1s to 24h selectable |
| Analysis software | 2000/XP/ Vista |
| Open input indication | "LO" appears on the LCD |
| Low battery indication | Empty battery symbol appears on the LCD |
| Power supply | 3.6V Lithium Battery |
| Battery life | 3 months (approximately) |
| Operating Temperature | -40 to 158°F, -40 to 70°C |
| Operating Humidity | 0 to 100% RH |
| LCD Operating Temperature | -13°F to 158°F (-25°C to 70°C) |
| Dimensions | 3.7x1.9x1.2" (94.4x48.9x31.2mm) |
| Weight | 3.2oz. (90.7g) |

| | Range | Accuracy |
|-------------------|------------------------------|----------|
| Relative Humidity | 0 to 20% and 80 to 100% | ±5.0% |
| | 20 to 40% and 60 to 80% | ±3.5% |
| | 40 to 60% | ±3.0% |
| Temperature | -40 to 14°F and 104 to 158°F | ±3.6°F |
| | 14 to 104°F | ±1.8°F |
| | -40 to -10°C and 40 to 70°C | ±2°C |
| | -10 to 40°C | ±1°C |

Copyright © 2009 Extech Instruments Corporation (a FLIR company)
All rights reserved including the right of reproduction in whole or in part in any form.